

ISO 9001

CERTIFIED

認証取得



MINOLTA

Inspection Certificate

(検査合格証)

PRODUCT : CHROMA METER

MODEL : CR-200, CR-210, CR-221, CR-231

SERIAL NO. : 73903006

This is to certify that the above product has been inspected with the results listed on the reverse and has been found to satisfy the inspection standards of this company.

上記製品は、裏面記載の項目について検査した結果、当社判定基準を満足していることを証明します。

Inspector K. Kobayashi

MINOLTA CO., LTD.

ITEMS INSPECTED		
APPEARANCE	Surface condition (scratches, stains); Assembly (looseness, misalignment, gaps); Printing	
FUNCTIONS	Display	Liquid crystal display, Viewing-angle adjustment
	Controls	Power switch, Keys
	Measuring Head	Measuring button, Ready lamp
	Program	Calibration, Measurement, Color space, Statistical calculations, Data list
	Printer	Data printout
	Input/Output	ΔE alarm output, Data input/output, Remote-control input, AC-adaptor input socket
PERFORMANCE	Repeatability	Standard deviation: ΔE^*ab within 0.07 Measurement conditions: White calibration plate measured 30 times at 10-second intervals

検査項目		
外観	傷・汚れ・隙間・段差・がた・表面処理・印刷	
機能	表示	液晶表示・視角調整
	キースイッチ	電源スイッチ・各種設定キー
	測定ヘッド	測定ボタン・レディランプ
	プログラム	校正・測定・表色・統計計算・データリスト
	プリンタ部	印字機能
	入出力	ΔE アラーム信号出力・データ信号入出力 リモート測定信号入力・ACアダプタ入力端子
性能	繰り返し性	標準偏差 ΔE^*ab : 0.07以内
	条件: 白色校正板を10秒間隔で30回測定し、当社評価プログラムにより判定	