

---

**Calibration Mark: 00630-PRI-18**

**Page 1/2**

This is to certify that the subject instrument was calibrated and tested at Acal BFi Germany GmbH on this date, using a working standard and other instruments whose accuracies are traceable to the National Institute of Standards and Technology (N.I.S.T.).

**Working Standard:**

Luminance Lamp: LRS-462 S/N 99406049  
Colorimetric coordinates:  $x = 0,4495$   $y = 0,4110$  CCT = 2856 K  
(for a Luminance: 1000 fl / 3427 cd/m<sup>2</sup>)

Irradiance Standard: 752-10U S/N 00102169  
Colorimetric coordinates:  $x = 0,4472$   $y = 0,4068$  CCT = 2856 K  
(for Illuminance: 73,55 fc / 791,7 lux)

Calibrated by Photo Research traceable to N.I.S.T.  
Primary Source of Calibration:  
Model No. 455-6 S/N: 10404374 N.I.S.T. Test No. 685/287969-16

The instrument was found to be within its rated accuracy of  $\pm 2$  % of Luminance/Radiance readings and within  $\pm 0,0015$  for CIE 1931 xy chromaticity coordinates when measuring a blackbody source at 2856 K. The spectral wavelength uncertainty (accuracy) of the instrument is less than  $\pm 2$  nm

**Article: Spectroradiometer**

**Model Name: PR-655**

**Reference Number: 682353**

**Serial Number: 65082002**

**Date of Calibration: 07.08.2018**

Measured values after calibration:

Aperture	Luminance of the source 1001,3 fl
	<b>MS-75</b>
1°	1,001 e+0023fl 2856 K x = 0,4494 y = 0,4110
Aperture	Luminance of the source 1001,5 fl
	<b>MS-2,5X</b>
1°	1,004 e+003 fl 2853 K x = 0,4496 y = 0,4110
Aperture	Luminance of the source 1001,5 fl
	<b>FP-655-2</b>
1°	1,002 e+003 fl 2857 K x = 0,4492 y = 0,4108

Performed by: Maxim Spevak

Date: 03.09.2018

Acal BFi Germany GmbH

Signed and Seal:

Oppelner Straße 5  
D-82194 Gröbenzell  
Tel. 0814276520-0, Fax: 190