

SP320R5 <input checked="" type="checkbox"/>	SP320R5D <input type="checkbox"/>	SP320R4 <input type="checkbox"/>	SP320R4D <input type="checkbox"/>
Model / model: SP320-114		SN : 30932004	
Kunde/customer: Universitat Polytechnica de Catalunya			
mit Zubehör with Accessory	ISP250-110 & ISP250-211 + OFG464 + PLG420		
	EOP146 + OFG444 + PLG440		
	LED434B + OFG464 + PLG420		
	TOP100 + OFG313-3MM +60mm Nikon		
Tra100-210+OFG313-3+PLG310			

Allgemein / general :		
	<input checked="" type="checkbox"/> JA / yes <input type="checkbox"/> Nein / no	Kommentar / comment:
Beschädigungen Gerät / <i>damages of the device</i>	<input type="checkbox"/>	Photo <input type="checkbox"/>
Starke Verschmutzung / <i>strong contamination</i>	<input type="checkbox"/>	Photo <input type="checkbox"/>
Gerät gereinigt / <i>device cleaned</i>	<input checked="" type="checkbox"/>	
Chopper-Funktion i.o / <i>Chopper function ok</i>	<input type="checkbox"/>	not available
Optical-Port -Funktion (LED leuchtet) / <i>optical port function(LED works)</i>	<input type="checkbox"/>	not available
PMT Kühlung vorhanden / <i>PMT cooling available</i>	<input type="checkbox"/>	not available
LED der PMT-Kühlung leuchtet / <i>LED of PMT cooling is working</i>	<input type="checkbox"/>	not available
LCD Anzeige i.o / <i>LCD panel ok</i>	<input checked="" type="checkbox"/>	Fehlercode/error Nr.
Gerät initialisiert / <i>device is initialising</i>	<input checked="" type="checkbox"/>	Power-On switch defective

Testmessungen / test measurements :			
Dauertest PMT-Empfindlichkeit <i>endurance test PMT sensitivity</i>	Dauer <i>runtime</i>	Messungen <i>measurements</i>	ok
durchgeführt / <i>done</i> <input checked="" type="checkbox"/>	180 min	240	<input checked="" type="checkbox"/>
Dauertest Mechanik/ <i>endurance test mechanics</i>	Dauer <i>runtime</i>	Messungen <i>measurements</i>	ok
durchgeführt / <i>done</i> <input checked="" type="checkbox"/>	180 min	90 (with 14 subranges)	<input checked="" type="checkbox"/>
Dunkelstrom komplett und selektiv <i>dark current complete and selective</i>	<input checked="" type="checkbox"/>	Detektor 1 PMT3 Wert / <i>value ok</i>	<input checked="" type="checkbox"/>
		detector 1 Detektor 2 InGaAs Wert / <i>value ok</i>	<input checked="" type="checkbox"/>
Messung mit der Tra100 <i>measurement with Tra100</i>	<input checked="" type="checkbox"/>	Gitter / <i>Grating</i> 1200UV Detektor 1 PMT3 Wert / <i>value ok</i>	<input type="checkbox"/>
		Gitter / <i>Grating</i> 600IR1 Detektor 2 InGaAs Wert / <i>value ok</i>	<input checked="" type="checkbox"/>

Durchführung von / <i>executed by:</i> UBI
Datum / <i>date:</i> 15.Jun. 2020

Messung mit der ACS530 <i>measurement with ACS530</i>	<input checked="" type="checkbox"/>	Gitter / Grating 1200UV	Detektor 1 <i>detector 1</i>	PMT3	Wert /value ok	<input type="checkbox"/>
		Gitter / Grating 600IR1	Detektor 2 <i>detector 2</i>	InGaAs	Wert /value ok	<input checked="" type="checkbox"/>
Messung mit der LED rot <i>measurement with LED red</i>	<input checked="" type="checkbox"/>	Gitter / Grating 1200UV	Detektor 1 <i>detector 1</i>	PMT3	Wert /value ok	<input type="checkbox"/>
		Gitter / Grating 600IR1	Detektor 2 <i>detector 2</i>	InGaAs	Wert /value ok	<input checked="" type="checkbox"/>
Wellenlängenüberprüfung mit da- zugehörigen Optik (mit Laser) <i>Wavelength check with corre- sponding optic (with Laser)</i>	<input checked="" type="checkbox"/>	Gitter / Grating 1200UV	Detektor 1 <i>detector 1</i>	PMT3	Wert /value ok	<input checked="" type="checkbox"/>
		Gitter / Grating 600IR1	Detektor 2 <i>detector 2</i>	InGaAs	Wert /value ok	<input checked="" type="checkbox"/>
Kantenfilterfaktoren überprüfen <i>check of the order sorting filter factor</i>		<input checked="" type="checkbox"/>	Abweichung i.o <i>abreviation o.k</i>			<input checked="" type="checkbox"/>
Dunkelmessung (nur SP320R4 /gekühlter NVIS) <i>dark measurement (only SP320R4/cooled NVIS)</i>		<input type="checkbox"/>	Offset i.o /offset ok.			<input type="checkbox"/>
PMT-Faktoren mit Tra100 / PMT factors with Tra100		<input checked="" type="checkbox"/>	Abweichung i.o / abreviation ok			<input checked="" type="checkbox"/>
Dauertest PMT-Rauschen HV4 (nur bei gek. PMT ;15 Min.) <i>endurance test PMT noise HV4 (only on cool. PMT;15 Min.)</i>		<input type="checkbox"/>	Drift des Rauschlevels i.o <i>drift of noise signal level ok</i>			<input type="checkbox"/>
Rauschscanmessung (PMT auf / on HV1) <input checked="" type="checkbox"/> <i>noise signal measurement</i>	Detektor 1/detector 1	PMT3	Gekühlt <i>cooled</i>	<input type="checkbox"/>	Bandbreite Rauschen i.o <i>bandwidth noise ok</i>	<input checked="" type="checkbox"/>
	Detektor 2/detector 2	InGaAs	Gekühlt <i>cooled</i>	<input checked="" type="checkbox"/>	Bandbreite Rauschen i.o <i>bandwidth noise ok</i>	<input checked="" type="checkbox"/>

Kontrolle aller mitgelieferter Zubehöre / checking of all received accessories			
Einkoppeloptik unbeschädigt <i>Optical probe not damaged</i>	Nr. 1 ISP250-110/211	<input checked="" type="checkbox"/>	Photo <input type="checkbox"/>
	Nr. 2 EOP146	<input checked="" type="checkbox"/>	Photo <input type="checkbox"/>
	Nr. 3 LED434B	<input checked="" type="checkbox"/>	Photo <input type="checkbox"/>
	Nr. 4 Tra100-210	<input checked="" type="checkbox"/>	Photo <input type="checkbox"/>
Faser i.o / fiber ok	Nr. 1 OFG464 (ISP250)	<input checked="" type="checkbox"/>	Photo <input type="checkbox"/>
	Nr. 2 OFG444 (EOP146)	<input checked="" type="checkbox"/>	Photo <input type="checkbox"/>
	Nr. 3 OFG464 (LED434B)	<input checked="" type="checkbox"/>	Photo <input type="checkbox"/>
	Nr. 4 OFG312-3 (Tra100-210)	<input checked="" type="checkbox"/>	Photo <input type="checkbox"/>
	Nr. 5 OFG313-3 (TOP100)	<input type="checkbox"/>	fiber crushed Photo <input checked="" type="checkbox"/>

Durchführung von / executed by: UBI

Datum / date: 15.Jun. 2020

TOP Objective vorhanden/ <i>TOP objective available</i>	<input checked="" type="checkbox"/>	Obj. 1 60mm (SN: 3188989)	Zustand / <i>condition</i> cover didn't works properly	Photo <input checked="" type="checkbox"/>
		Obj. 2 (SN:)	Zustand / <i>condition</i>	Photo <input type="checkbox"/>
TOP100 vorhanden / <i>available</i>	<input checked="" type="checkbox"/>		Zustand / <i>condition</i> o.k.	Photo <input type="checkbox"/>
Blendenumschaltung i.o <i>aperture switching ok</i>	<input checked="" type="checkbox"/>			
View/Measure Umschaltung i.o <i>View/Measure switching ok</i>	<input checked="" type="checkbox"/>			
Schielen i.o <i>peering ok</i>	<input checked="" type="checkbox"/>			
TOP-Magnet i.o / ok	<input checked="" type="checkbox"/>			
TOP-Distanzaufkleber vorhanden <i>TOP-distance label available</i>	<input checked="" type="checkbox"/>	Wenn nein, wurde er nachgerüstet / <i>if not ,has it been upgraded</i>		<input type="checkbox"/>
TOP Transportsicherung vorhanden <i>TOP transport safety device available</i>	<input checked="" type="checkbox"/>	Wenn nein, wurde es nachgerüstet / <i>if not ,has it been upgraded</i>		<input type="checkbox"/>
TOP200 vorhanden / <i>available</i>	<input type="checkbox"/>	SN.	Zustand / <i>condition</i>	Photo <input type="checkbox"/>
Blendenumschaltung i.o <i>aperture switching ok</i>	<input type="checkbox"/>			
Blendenrad leichtgängig <i>aperture wheel free-moving</i>	<input type="checkbox"/>			
Overlay-Verschiebung getestet <i>Overlay shift tested</i>	<input type="checkbox"/>			
Weißabgleich getestet	<input type="checkbox"/>			
White balance tested	<input type="checkbox"/>			
Vignettierung getestet <i>Vignetting tested</i>	<input type="checkbox"/>			
Blenden sind sauber und unbeschädigt <i>apertures not contaminated and damaged</i>	<input type="checkbox"/>	Blende / <i>aperture 1</i>	Zustand / <i>condition</i>	Photo <input type="checkbox"/>
	<input type="checkbox"/>	Blende / <i>aperture 2</i>	Zustand / <i>condition</i>	Photo <input type="checkbox"/>
	<input type="checkbox"/>	Blende / <i>aperture 3</i>	Zustand / <i>condition</i>	Photo <input type="checkbox"/>
	<input type="checkbox"/>	Blende / <i>aperture 4</i>	Zustand / <i>condition</i>	Photo <input type="checkbox"/>
	<input type="checkbox"/>	Blende / <i>aperture 5</i>	Zustand / <i>condition</i>	Photo <input type="checkbox"/>
	<input type="checkbox"/>	Blende / <i>aperture 6</i>	Zustand / <i>condition</i>	Photo <input type="checkbox"/>
	<input type="checkbox"/>	Offset	Zustand / <i>condition</i>	Photo <input type="checkbox"/>
TOP150 vorhanden / <i>available</i>	<input type="checkbox"/>		Zustand / <i>condition</i>	Photo <input type="checkbox"/>
Blende sauber und unbeschädigt <i>aperture not contaminated or damaged</i>	<input type="checkbox"/>			

 Durchführung von / *executed by*: UBI

 Datum / *date*: 15.Jun. 2020

ISP250-110 vorhanden / available <input checked="" type="checkbox"/>	SN. 05692007	Zustand / condition o.k. Photo <input type="checkbox"/>
Beschichtung i.o / coating ok	<input checked="" type="checkbox"/>	
Hilfslichtquelle i.o	<input checked="" type="checkbox"/>	
<i>auxiliary light source ok</i>		
Keine Fluoreszenz / no fluorescence (ISP75...2000)	<input type="checkbox"/>	

Durchführung von / executed by: UBI

Datum / date: 15.Jun. 2020

RMA: 31885	SN: 30932004	Gerät / device: SP320-114
Wareneingangsdatum / receipt: 04.June. 2020		
Kunde / customer		
Universität Polytechnica de Catalunya		
CD6-Centre Desemvolupament de Sensors		
c/ Rambla Sant Nebridi ,10		
08222 Terassa-Barcelona ,Spain		
Grund der Einsendung / reason for return		
Repair (Tra100-210→ Lamp not stable) / Recalibration		
Power switch defective		
<input type="checkbox"/> Garantie / warranty <input type="checkbox"/> Kulanz / goodwill <input checked="" type="checkbox"/> Rechnung / on charge		

Fehlfunktion / malfunction
1) TOP100: OFG313 fiber damaged →“Aperture 3“ image is showing blocked/shadowed light path zone
2) Tra100-210 Lamp: during test with InGaAs (17 hours) → measurement abbreviation (radiometric integral) max. 0,15%
3) Sensitivity of the PMT3 detector decreased about 9 % relating to 2014 →this is normally for the PMT3 detector because it is aging temporally depending of diverse factors
4) Power switch defective → sporadically without function

Durchgeführte Arbeiten / performed work
SRV-0302 General instrument check
To 1) replaced the damaged OFG313 fiber
To 2) replaced the lamp (as a precaution →lamp is 16 years old)
To 4) replace the power switch
Calibrations →as ordered

Funktionstest nach der Reparatur / functional test after repair
Endurance test of mechanics & electronics
VDE safety test

Empfehlungen zur Fehlervermeidung / recommendations for error prevention
To 3) On the “SP320 user manual “we recommend following calibration interval: The calibration interval is normally 1 year. Under difficult operating conditions, i.e. if the SPECTRO 320 (D) is subject to contamination (room air) or severe temperature variations, calibration must be carried out at shorter intervals. Detectors are subject to an aging process if they are used for long periods. This must be taken into account and in the case of UV detectors in particular. Shorter calibration intervals are also recommended here.

Platz für Messergebnisse oder Fotos / space for measurement results or photos:


Ausgetauschte Teile werden bei Instrument Systems gelagert und nach einem Monat fachgerecht entsorgt.
Defect parts will be kept at Instrument Systems and disposed of after one month.

Durchführung von / executed by: UBI
Datum / date: 02.07. 2020

TEST CERTIFICATE

Certificate No.	CAL-200-20-015	
Instrument	Spectroradiometer	
Manufacturer	Instrument Systems Optische Messtechnik GmbH Kastenbauerstr. 2 81677 München	
Instrument Type / Serial Number	SP320 -114	SN: 30932004
Accessories / Serial Numbers	LED-434	The serial number of the instrument is provided on the optical fiber to ensure the proper assignment of the external optical probe to the instrument.
	OFG-464	
	PLG-420	
Type of Test	Determination of relative spectral irradiance (λ) in the wavelength range from 800 nm to 1700 nm in conformity with CIE63:1984	
Date of Test	26. Jun. 2020	
Customer	Universitat Polytécnica de Catalunya CD6 - Centre de Desenvolupament de Sensors c/ Rambla de Sant Nebridi 10 08222 Terrassa - Barcelona , Spain	
Purchase Order No.	SAR001782	

Instrument Systems Optische Messtechnik GmbH

Date	Prepared by	Approved by
02.07.2020		
	U. Binder Test Lab and Service Engineer	M. Steinbach Test Lab and Service Engineer

1. Instrument Description

The above named instrument under test is an optical spectroradiometer. The light, coupled into the spectrometer via an optical fiber, is split into its spectral components by means of a grating and then measured by a suitable detector. The measurement of the irradiance is carried out using an optical probe, which is fixed to the other end of the optical fiber.

2. Test Procedure

a) Wavelength accuracy:

The wavelength scale of the spectrometer is established and verified using emissions of spectral lamps and lasers. These emit one or more spectral lines of known wavelengths listed in the NIST Atomic Spectra Database. Lamps used in the test procedure include Hg, Ar and Xe lines. Other wavelengths are included using HeNe gas lasers. The line wavelengths of the test sources used for the wavelength calibration procedure are physical standards and do not require further traceability. In the infrared wavelength range above 1000 nm Nd based solid state lasers are also used. The wavelength of the lasers is determined using a wavemeter or reference spectrometer.

b) Spectral sensitivity:

The spectral sensitivity of the spectrometer is determined using an appropriate light source with a relative spectral distribution of radiation $S(\lambda)$ that is known to lie within the spectrometer's wavelength range. For the visible and infrared wavelength range, a halogen lamp is used with a spectral distribution similar to a Planckian radiator with a color temperature of approximately 3100 K.

The measuring area of the optical probe is positioned centrally and perpendicularly to the optical axis of the defined radiation direction of the lamp.

The following configuration files must be installed in the spectrometer and/or the measuring software to ensure proper test results:

309320L1.ini	Configuration file containing information about the wavelength scale, the absolute sensitivity and the hardware parameters of the spectrometer.
309320L1.isc	Calibration file containing the correction of the spectral sensitivity with detector InGaAs.

Certificate No.: CAL-200-20-015

3. Measurement conditions and traceability

Used reference standards and their traceability:

Working Standard	Internal Calibration Reference	Date of Calibration	Traceable to	External Calibration Reference
1000W halogen lamp BN-9101-618	CEQ-252-20-006	Jan. 2020	Reference standard BN-9101-711	40020-18-PTB

Ambient conditions:

Ambient conditions	
Temperature	23 °C ± 3°C
Humidity	53% ± 5% relative humidity

Spectrometer settings during irradiance measurement:

Detector :	InGaAs	Grating :	600 g/mm IR1
Wavelength range :	800 - 1700	Bandpass :	2 nm
Scan speed:	60 ms / nm	Density filter:	none
Step width:	1 nm	Cut filter :	yes
Averaging:	2		

The warm-up time of the spectrometer was 2 hours. The spectrometer SP320-114 and all accessories listed on page 1 were in good order and condition during the entire calibration process.

Certificate No.: CAL-200-20-015

4. Measurement results

a) Wavelength

Table of measured lines:

Nominal value [nm]	Measured value [nm]
881.94	881.94
979.97	980.01
1064.27	1064.3
1152.27	1152.28
1531.93	1531.94

The measurement tolerance of the wavelength measurement is +/-0.2 nm for all wavelengths.

b) Relative spectral irradiance

For the testing of the measurement accuracy of the spectroradiometer for relative irradiance the measurement parameters and working standard mentioned in point 3 were used.

Table of measured relative irradiance values:

Wavelength [nm]	Relative irradiance Nominal value	Relative irradiance Measured value
800	0.949613	0.948681
900	0.998345	0.997989
1000	0.981498	0.981285
1100	0.926806	0.926648
1200	0.845145	0.845217
1300	0.757005	0.756981
1400	0.675115	0.674837
1500	0.593055	0.592623
1600	0.517725	0.517466

Certificate No.: CAL-200-20-015

5. Notes

- a) Any mechanical modifications to the tested instrument and the accessories listed on page 1 influence the sensitivity of the instrument.
- b) Instrument Systems certifies that all reference standards and measuring devices have been used within the valid scope of application during the entire test process. The test objects described on page 1 were in good order and condition during the test procedure.
- c) This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.
- d) The results indicated in this certificate refer solely to the instruments tested in the laboratory.
- e) Strong UV-radiation can cause a significant aging of optical components. The measurement results can be influenced by that aging.
- f) For measurements with the largest bandpass (10nm with 600g/mm grating) the spectral sensitivity needs to be corrected. That is done using the calibration files „LED434-B, InGaAs calibrated on Bandpass 10nm “ (309320L2.ini /309320L2.isc).

End of certificate



TEST CERTIFICATE

Certificate No. CAL-203-20-014

Instrument Spectroradiometer

Manufacturer Instrument Systems Optische Messtechnik GmbH
Kastebauerstr. 2
81677 München

Instrument Type / Serial Number SP320 -114 SN: 30932004

Accessories / Serial Numbers EOP-146
OFG-444
PLG-444

The serial number of the instrument is provided on the optical fiber to ensure the proper assignment of the external optical probe to the instrument.

Type of Test Determination of spectral irradiance $E_e(\lambda)$ in the wavelength range from 280 nm to 1700 nm in conformity with CIE63:1984

Date of Test 25. Jun. 2020

Customer Universitat Polytécnica de Catalunya
CD6 - Centre de Desenvolupament de Sensors
c/ Rambla de Sant Nebridi, 10
08222 Terrassa - Barcelona
Spain

Purchase Order No. SAR001782

Instrument Systems Optische Messtechnik GmbH

Date
02.07.2020

Prepared by

U. Binder
Test Lab and Service Engineer

Approved by

M. Steinbach
Test Lab and Service Engineer

1. Instrument Description

The above named instrument under test is an optical spectroradiometer. The light, coupled into the spectrometer via an optical fiber, is split into its spectral components by means of a grating and then measured by a suitable detector. The measurement of the irradiance is carried out using an optical probe, which is fixed to the other end of the optical fiber.

2. Test Procedure

a) Wavelength accuracy:

The wavelength scale of the spectrometer is established and verified using emissions of spectral lamps and lasers. These emit one or more spectral lines of known wavelengths listed in the NIST Atomic Spectra Database. Lamps used in the test procedure include Hg, Ar and Xe lines. Other wavelengths are included using HeNe gas lasers. The line wavelengths of the test sources used for the wavelength calibration procedure are physical standards and do not require further traceability. In the infrared wavelength range above 1000 nm Nd based solid state lasers are also used. The wavelength of the lasers is determined using a wavemeter or reference spectrometer.

b) Spectral sensitivity:

The spectral sensitivity of the spectrometer is determined using an appropriate light source with a relative spectral distribution of radiation $S(\lambda)$ that is known to lie within the spectrometer's wavelength range. For the visible and infrared wavelength range, a halogen lamp is used with a spectral distribution similar to a Planckian radiator with a color temperature of approximately 3100 K. For the ultraviolet part of the spectrum, a deuterium lamp is applied. The measuring area of the optical probe is positioned centrally and perpendicularly to the optical axis of the defined radiation direction of the lamp.

c) Irradiance measurement:

The test procedure is based on the following standard: CIE Commission Internationale de L'Eclairage, "The spectroradiometric measurement of light sources", Publication CIE 63:1984.

The irradiance measurement is carried out using an appropriate light source with a known spectral irradiance $E(\lambda)$ for a specified radiation direction and a predefined distance. This light source homogeneously radiates light onto the optical probe. The distance between the front face of the optical probe and the reference plane of the lamp relevant for the measured spectral irradiance is 70 cm.

The following configuration files must be installed in the spectrometer and/or the measuring software to ensure proper test results:

309320E1.ini	Configuration file containing information about the wavelength scale, the absolute sensitivity and the hardware parameters of the spectrometer.
309320E1.isc	Calibration file containing the correction of the spectral sensitivity with detector PMT3 and detector InGaAs.

Certificate No.: CAL-203-20-014

3. Measurement conditions and traceability

Used reference standards and their traceability:

Working Standard	Internal Calibration Reference	Date of Calibration	Traceable to	External Calibration Reference
1000W halogen lamp BN-9101-618	CEQ-252-20-006	Jan. 2020	Reference standard BN-9101-711	40020-18-PTB
Deuterium lamp AN5083	CEQ-262-19-005	Nov. 2019	Reference standard DLS-IS-L#01	40022-19-PTB

Ambient conditions:

Ambient conditions	
Temperature	23 °C ± 3°C
Humidity	55% ± 5% relative humidity

Spectrometer settings during irradiance measurement:

Detector 1 :	PMT3	Grating 1 :	1200 g/mm UV
Detector 2 :	InGaAs	Grating 2 :	600 g/mm IR1
Wavelength range 1:	280 – 930 nm	Bandpass 1 :	2.5 nm
Wavelength range 2:	800 - 1700	Bandpass 2 :	5 nm
Scan speed:	60 ms / nm	Density filter:	none
Step width:	1 nm	Cut filter :	yes
Averaging:	2		
PMT high voltage:	3	PMT cooling:	none

The warm-up time of the spectrometer was 2 hours. The spectrometer SP320-114 and all accessories listed on page 1 were in good order and condition during the entire calibration process.

Certificate No.: CAL-203-20-014

4. Measurement results

a) Wavelength

Table of measured lines:

Nominal value [nm]	Measured value [nm]
435.83	435.88
546.08	546.13
632.82	632.88
881.94	881.91
979.97	979.98
1064.27	1064.3
1152.27	1152.37
1309.68	1309.69
1531.93	1531.91

The measurement tolerance of the wavelength measurement is +/-0.2 nm for all wavelengths.

b) Irradiance

For the testing of the measurement accuracy of the spectroradiometer for irradiance the measurement parameters and working standard mentioned in point 3 were used.

Table of measured irradiance values

Wavelength [nm]	Irradiance Nominal value [W/(m ² nm)]	Irradiance Measured value [W/(m ² nm)]
400	9.300E-03	9.263E-03
500	3.186E-02	3.177E-02
600	6.110E-02	6.115E-02
700	8.576E-02	8.576E-02
800	1.007E-01	1.007E-01
900	1.059E-01	1.061E-01
1000	1.041E-01	1.042E-01
1200	8.964E-02	8.966E-02
1400	7.161E-02	7.165E-02
1600	5.443E-02	5.495E-02

Certificate No.: CAL-203-20-014

The measurement uncertainty of the spectral irradiance depends on the wavelength:

Wavelength λ in [nm]	Relative expanded uncertainty
$280 \leq \lambda < 300$	9%
$300 \leq \lambda < 360$	7%
$360 \leq \lambda < 400$	4%
$400 \leq \lambda < 800$	3.5%
$800 \leq \lambda < 850$	4%
$850 \leq \lambda < 950$	5%
$950 \leq \lambda < 1650$	4.5%
$1650 \leq \lambda < 1700$	6%

The specified relative expanded uncertainty of measurement corresponds to twice the standard deviation ($k=2$). For a normal distribution, this means that 95 per cent of the measured values lie within the indicated measurement uncertainty interval.

5. Notes

- a) Any mechanical modifications to the tested instrument and the accessories listed on page 1 influence the sensitivity of the instrument.
- b) Instrument Systems certifies that all reference standards and measuring devices have been used within the valid scope of application during the entire test process. The test objects described on page 1 were in good order and condition during the test procedure.
- c) This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.
- d) The results indicated in this certificate refer solely to the instruments tested in the laboratory.
- e) Strong UV-radiation can cause a significant aging of optical components. The measurement results can be influenced by that aging.
- f) For measurements with the largest bandpass (10nm with 600g/mm grating) the spectral sensitivity needs to be corrected. That is done using the calibration files „EOP146, InGaAs calibrated on Bandpass 10nm“ (309320E2.ini /309320E2.isc).

End of certificate



TEST CERTIFICATE

Certificate No. CAL-221-20-002

Instrument Spectroradiometer

Manufacturer Instrument Systems Optische Messtechnik GmbH
Kastenbauerstr. 2
81677 München

Instrument Type / Serial Number SP320 -114 SN: 30932004

Accessories / Serial Numbers LED-434 with LED-439
OFG-464
PLG-420

The serial number of the instrument is provided on the optical fiber to ensure the proper assignment of the external optical probe to the instrument.

Type of Test Test of averaged LED intensity $I_{LED,B}$ [cd] of LEDs in conformity with CIE127:2007.

Date of Test 26. Jun. 2020

Customer Universitat Polytécnica de Catalunya
CD6 - Centre de Desenvolupament de Sensors
c/ Rambla de Sant Nebridi
10 08222 Terrassa - Barcelona , Spain

Purchase Order No. SAR001782

Instrument Systems Optische Messtechnik GmbH

Date

02.07.2020

Prepared by

U. Binder
Test Lab and Service Engineer

Approved by

M. Steinbach
Test Lab and Service Engineer

1. Instrument Description

The above named instrument under test is an optical spectroradiometer. The light, coupled into the spectrometer via an optical fiber, is split into its spectral components by means of a grating and then measured by a suitable detector. The measurement of the Averaged LED intensity of LEDs is carried out using a detector head which is fixed to the other end of the optical fiber.

The detector head comprises a plane, circular diffuser with an area of 100 mm². A spacer tube placed in front of the diffuser and a suitable LED test socket are used to ensure the correct positioning of the LED, centrally and perpendicularly to the diffuser and with a distance of 100 mm ($I_{LED,B}$) and 316 mm ($I_{LED,A}$), respectively, to the LED's light emitting area.

2. Test Procedure

a) Wavelength accuracy:

The wavelength scale of the spectrometer is established and verified using emissions of spectral lamps and lasers. These emit one or more spectral lines of known wavelengths listed in the NIST Atomic Spectra Database. Lamps used in the test procedure include Hg, Ar and Xe lines. Other wavelengths are included using HeNe gas lasers. The line wavelengths of the test sources used for the wavelength calibration procedure are physical standards and do not require further traceability. In the infrared wavelength range above 1000 nm Nd based solid state lasers are also used. The wavelength of the lasers is determined using a wavemeter or reference spectrometer.

b) Spectral sensitivity:

The spectral sensitivity of the spectrometer is determined using one or more appropriate light sources with a relative spectral distribution of radiation $S(\lambda)$ that is known to lie within the spectrometer's wavelength range. A halogen lamp with a spectral distribution similar to a Planckian radiator with a colour temperature of approximately 3100 K is used for the visible (and infrared) wavelength range between 360nm and 930nm. For the ultraviolet part of the spectrum, a deuterium lamp is used in the wavelength range between 200nm and 360nm. The radiation of the lamp is coupled into the detector head, centrally and perpendicularly to the diffuser. The spacer tube for the positioning of the LED is removed during this procedure.

c) Averaged LED intensity of LEDs:

The test procedure is based on the following standard: CIE Commission Internationale de L'Eclairage, "Measurement of LEDs", Publication CIE 127:2007 2nd edition.

The measurement is carried out using high-power LED calibration standards with calibrated averaged LED intensity values. The LED calibration standards are operated with a constant current and at a stabilized temperature. The light of the LED is emitted from the front face of a metal tube with a diameter of 25 mm that is mounted on the top of the housing. The LED calibration standards are mounted to an appropriate test socket and positioned in the required distance to the detector head according to CIE127:2007 $I_{LED,B}$.

The following configuration files must be installed in the spectrometer and/or the measuring software to ensure proper test results:

309320L1.ini	Configuration file containing information about the wavelength scale, the absolute sensitivity and the hardware parameters of the spectrometer.
309320L1.isc	Calibration file containing the correction of the spectral sensitivity with detector PMT3.

Certificate No.: CAL-221-20-002

3. Measurement conditions and traceability

Used reference standards and their traceability:

Working Standard	Internal Calibration Reference	Date of Calibration	Traceable to	External Calibration Reference
1000W halogen lamp BN-9101-618	CEQ-252-20-006	Jan. 2020	Reference standard BN-9101-711	40020-18-PTB
Deuterium lamp AN5083	CEQ-262-19-005	Nov. 2019	Reference standard DLS-IS-L#01	40022-19-PTB
LED calibration standard white, ACS-570-1 SN: 007357018	LED-19-017	Sep. 2019	Reference standards: - ACS570-1 SN:10008751 - ACS570-1 SN:10008851	47237 PTB 17 47233 PTB 17
LED calibration standard blue, ACS-570-3 SN: 008457018	LED-19-019	Sep. 2019	Reference standards: - ACS570-3 SN:10051551 - ACS570-3 SN:10051651	47238 PTB 17 47234 PTB 17
LED calibration standard green, ACS-570-5 SN: 007557018	LED-19-021	Sep. 2019	Reference standards: - ACS570-5 SN:000457017 - ACS570-5 SN:10052051	47239 PTB 17 47235 PTB 17
LED calibration standard red, ACS-570-7 SN: 007757018	LED-19-023	Sep. 2019	Reference standards: - ACS570-7 SN:000957017 - ACS570-7 SN:10061451	47240 PTB 17 47236 PTB 17

Ambient conditions:

Temperature 23 °C +/- 3 °C

Humidity 54 % relative humidity

Spectrometer settings during averaged LED intensity measurement:

Detector:	PMT3	Averaging:	2
Grating:	1200 g/mm UV	Step width:	1 nm
Wavelength range:	200 – 930 nm	Density filter:	none
Bandpass:	1 nm	Cut filter :	yes
Scan speed:	60 ms / nm	PMT high voltage:	3
PMT cooling:	none		

The warm-up time of the spectrometer was 2 hours. The spectrometer SP320-114 and all accessories listed on page 1 were in good order and condition during the entire calibration process.

Certificate No.: CAL-221-20-002

4. Measurement results

a) Wavelength

Table of measured lines:

Nominal value [nm]	Measured value [nm]
253.65	253.66
435.84	435.85
546.08	546.09
632.82	632.84
881.94	881.94

The measurement tolerance of the wavelength measurement is 0.2 nm for all wavelengths.

b) Averaged LED intensity of LEDs

For the testing of the measurement accuracy of the spectroradiometer for averaged LED intensity the measurement parameters and working standards mentioned in point 3 were used.

Table of results:

LED Serial number	Dominant wavelength [nm] / correlated color temperature [K]	Averaged LED intensity [cd]	Averaged LED intensity [cd]
		Nominal value	Measured value
ACS570-1_007357018	5905	18.83	18.83
ACS570-3_008457018	481.34	8.647	8.797
ACS570-5_007557018	528.31	18.58	18.54
ACS570-7_007757018	622.34	9.701	9.702

The values of dominant wavelength and correlated color temperature are not subject to the test procedure and specified for information only.

The measurement uncertainty of the averaged LED intensity depends on the color of the LED:

Color	relative expanded uncertainty
Blue	4.5%
White	3%
Green	3%
Red	4.5%

The specified relative expanded uncertainty of measurement corresponds to twice the standard deviation ($k=2$). For a normal distribution, this means that 95 per cent of the measured values lie within the indicated measurement uncertainty interval.

5. Notes

- a) Any mechanical modifications to the tested instrument and the accessories listed on page 1 influence the sensitivity of the instrument.
- b) Instrument Systems certifies that all reference standards and measuring devices have been used within the valid scope of application during the entire test process. The test objects described on page 1 were in good order and condition during the test procedure.
- c) This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.
- d) The results indicated in this certificate refer solely to the instruments tested in the laboratory.
- e) The accessory LED-439 which is used for this testing is not the property of the customer and will not be delivered with this purchase order.

End of certificate



TEST CERTIFICATE

Certificate No. CAL-230-20-013

Instrument Spectroradiometer

Manufacturer Instrument Systems Optische Messtechnik GmbH
Kastenbauerstr. 2
81677 München

Instrument Type / Serial Number SP320 -114 SN: 30932004

Accessories / Serial Numbers TOP100-100 SN: --
Nikon AF 60mm lens SN: 3188989
TOP100-101 The serial number of the instrument is provided on the optical fiber to ensure the proper assignment of the external optical probe to the instrument.

Type of Test Determination of luminance L_v and spectral radiance $L_e(\lambda)$ in the wavelength range from 380 nm to 930 nm in conformity with CIE63:1984, ISO 23539:2005 und DIN 13032-1:2012.

Date of Test 30. Jun. 2020

Customer Universitat Polytécnica de Catalunya
CD6 - Centre de Desenvolupament de Sensors
c/ Rambla de Sant Nebridi
10 08222 Terrassa - Barcelona , Spain

Purchase Order No. SAR001782

Instrument Systems Optische Messtechnik GmbH

Date
02.07.2020

Prepared by

U. Binder
Test Lab and Service Engineer

Approved by

M. Steinbach
Test Lab and Service Engineer

1. Instrument Description

The above named instrument under test is an optical spectroradiometer. The light, coupled into the spectrometer via an optical fiber, is split into its spectral components by means of a grating and then measured by a suitable detector. The measurement of the radiance is carried out using an optical probe, which is fixed to the other end of the optical fiber.

2. Test Procedure

a) Wavelength accuracy:

The wavelength scale of the spectrometer is established and verified using emissions of spectral lamps and lasers. These emit one or more spectral lines of known wavelengths listed in the NIST Atomic Spectra Database. Lamps used in the test procedure include Hg, Ar and Xe lines. Other wavelengths are included using HeNe gas lasers. The line wavelengths of the test sources used for the wavelength calibration procedure are physical standards and do not require further traceability. In the infrared wavelength range above 1000 nm Nd based solid state lasers are also used. The wavelength of the lasers is determined using a wavemeter or reference spectrometer.

b) Spectral sensitivity:

The spectral sensitivity of the spectrometer is determined using a standard lamp for radiance with a relative spectral distribution of radiation $S(\lambda)$ that is known to lie within the spectrometer's wavelength range. The spectral distribution is similar to a Planckian radiator with a color temperature of approximately 2856 K. The optical axis of the telescopic optical probe lies perpendicular to the luminous area of the radiance standard. This area is in the focus of the objective lens.

c) Luminance and radiance:

The test procedure is based on the following standards: CIE Commission Internationale de L'Eclairage, "The spectroradiometric measurement of light sources", Publication CIE 63:1984, and "Photometry — The CIE system of physical photometry", Publication ISO 23539:2005 (CIE S010:2004).

The luminance is determined focusing the objective lens of the optical probe onto the luminous area of a working standard with known average luminance and radiance values. The distance between the front plane of the housing of the telescopic optical probe and the luminous area of the standard lamp is 500mm.

The following configuration files must be installed in the spectrometer and/or the measuring software to ensure proper test results:

309320T1.ini	Configuration file containing information about the wavelength scale, the absolute sensitivity and the hardware parameters of the spectrometer.
309320T1.isc	Calibration file containing the correction of the spectral sensitivity with detector PMT3.

Certificate No.: CAL-230-20-013

3. Measurement conditions and traceability

Measurements were taken under conditions in conformity with DIN EN 13032-1:2012.

Used reference standards and their traceability:

Working standards	Internal calibration reference	Date of calibration	Traceable to	External calibration reference
Luminance and radiance standard Hoffman Engineering LS-65-8E #HEC7936	CEQ-274-20-011	19.06.2020	- Reference standard FEL-1000W BN-9101-712 - Reflectance standard SRS-99-020 no. 4029 - Luminance standard LN3 SN: 03B202	40021-18-PTB PTB 44190/17 40059 PTB 12

Ambient conditions:

Temperature 23 °C +/- 3 °C

Humidity 57 % relative humidity

Spectrometer settings during radiance measurement:

Detector:	PMT3	Averaging:	2
Grating:	1200 l/mm UV	Step width:	1 nm
Wavelength range:	380 – 930 nm	Density filter:	none
Bandpass:	5 nm	Cut filter :	yes
Scan speed:	60 ms / nm	TOP100 aperture size:	2
PMT high voltage:	3	“ f number of lens:	2.8
PMT cooling:	none	“ focal length of lens:	60 mm

The warm-up time of the spectrometer was 3 hours.

Certificate No.: CAL-230-20-013

4. Measurement results

a) Wavelength

Table of measured lines:

Nominal value [nm]	Measured value [nm]
435.84	435.88
546.08	546.14
632.82	632.90
881.94	881.90

The measurement tolerance of the wavelength measurement is 0.2 nm for all wavelengths.

b) Luminance and radiance

For the testing of the measurement accuracy of the spectroradiometer for spectral radiance and luminance the measurement parameters and working standard mentioned in point 3 were used.

Table of results:

Wavelength [nm]	Radiance L_e Nominal value [W/(cm ² sr nm)]	Radiance L_e Measured value [W/(cm ² sr nm)]
	400	7.939E-08
500	4.003E-07	4.004E-07
600	8.638E-07	8.645E-07
700	1.288E-06	1.285E-06
800	1.589E-06	1.592E-06
900	1.751E-06	1.754E-06
Luminance L_v	Nominal value [cd/m ²]	Measured value [cd/m ²]
	494.5	494.6

Certificate Nr.: CAL-230-20-013

The measurement uncertainty of the spectral radiance depends on the wavelength:

Wavelength λ in [nm]	Relative expanded uncertainty
$380 \leq \lambda < 400$	5%
$400 \leq \lambda < 500$	4%
$500 \leq \lambda < 850$	3.5%
$850 \leq \lambda < 930$	4.5%

The measurement uncertainty for luminance L_v relates to the spectra of halogen lamps with a color temperature of approximately 2856 K and amounts to:

Luminance L_v	Relative expanded uncertainty
	3.5%

The specified relative expanded uncertainty of measurement corresponds to twice the standard deviation ($k=2$). For a normal distribution, this means that 95 per cent of the measured values lie within the indicated measurement uncertainty interval.

5. Notes

- Any mechanical modifications to the tested instrument and the accessories listed on page 1 influence the sensitivity of the instrument.
- Instrument Systems certifies that all reference standards and measuring devices have been used within the valid scope of application during the entire test process. The test objects described on page 1 were in good order and condition during the test procedure.
- This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.
- The results indicated in this certificate refer solely to the instruments tested in the laboratory.

End of certificate



TEST CERTIFICATE

Certificate No. CAL-241-20-001

Instrument Spectroradiometer

Manufacturer Instrument Systems Optische Messtechnik GmbH
Kastenbauerstr. 2
81677 München

Instrument Type / Serial Number SP320 -114 SN: 30932004

Accessories / Serial Numbers ISP-250 SN: 05692007
OFG-464
PLG-420

The serial number of the instrument is provided on the optical fiber to ensure the proper assignment of the external optical probe to the instrument.

Type of Test Test of luminous flux [lm] of LEDs in conformity with CIE127 - 2007


Date of Test 29. Jun. 2020


Customer Universitat Polytécnica de Catalunya
CD6 - Centre de Desenvolupament de Sensors
c/ Rambla de Sant Nebridi
10 08222 Terrassa - Barcelona , Spain

Purchase Order No. SAR001782

Instrument Systems Optische Messtechnik GmbH

Date
02.07.2020

Prepared by

U. Binder
Test Lab and Service Engineer

Approved by

M. Steinbach
Test Lab and Service Engineer

1. Instrument Description

The above named instrument under test is an optical spectroradiometer. The light, coupled into the spectrometer via an optical fiber, is split into its spectral components with a diffraction grating and then measured by a suitable detector. The measurement of the luminous flux of LEDs is carried out using an integrating sphere which is fixed to the other end of the optical fiber.

2. Test Procedure

a) Wavelength accuracy:

The wavelength scale of the spectrometer is established and verified using emissions of spectral lamps and lasers. These emit one or more spectral lines of known wavelengths listed in the NIST Atomic Spectra Database. Lamps used in the test procedure include Hg, Ar and Xe lines. Other wavelengths are included using HeNe gas lasers. The line wavelengths of the test sources used for the wavelength calibration procedure are physical standards and do not require further traceability. In the infrared wavelength range above 1000 nm Nd based solid state lasers are also used. The wavelength of the lasers is determined using a wavemeter or reference spectrometer.

b) Spectral sensitivity:

The spectral sensitivity of the spectrometer is determined using one or more appropriate light sources with a relative spectral distribution of radiation $S(\lambda)$ that is known to lie within the spectrometer's wavelength range. A halogen lamp with a spectral distribution similar to a Planckian radiator with a colour temperature of approximately 3100 K is used for the visible (and infrared) wavelength range between 360nm and 930nm. For the ultraviolet part of the spectrum, a deuterium lamp is used in the wavelength range between 240nm and 360nm. The radiation of the lamp is coupled in via the side port of the integrating sphere. Apart from the auxiliary lamp and the detector including the corresponding shutters, there are no further accessories inside the sphere.

c) Luminous flux of LEDs:

The test procedure is based on the following standard: CIE Commission Internationale de L'Eclairage, "Measurement of LEDs", Publication CIE 127:2007 2nd edition.

The measurement is carried out using high-power LED calibration standards with calibrated luminous flux values. The LED calibration standards are operated with a constant current and at a stabilized temperature. The light of the LED is emitted from the front face of a metal tube with a diameter of 25 mm that is mounted on the top of the housing. The luminous flux is measured using an integrating sphere. The LED is positioned at the edge of the integrating sphere, with the light emitting area extending approximately 2 mm into the sphere. The diameter of the aperture of the side port is 10 mm. The mechanical design of the LED prevents any backward-directed radiation.

For calibration the adapter plate ISP250-211 is mounted at the port of the integrating sphere to insert the LED test socket.

The following configuration files must be installed in the spectrometer and/or the measuring software to ensure proper test results:

30932012.ini	Configuration file containing information about the wavelength scale, the absolute sensitivity and the hardware parameters of the spectrometer.
30932012.isc	Calibration file containing the correction of the spectral sensitivity with detector PMT3.

Certificate No.: CAL-241-20-001

3. Measurement conditions and traceability

Used reference standards and their traceability:

Working Standard	Internal Calibration Reference	Date of Calibration	Traceable to	External Calibration Reference
1000W halogen lamp BN-9101-618	CEQ-252-20-006	Jan. 2020	Reference standard BN-9101-711	40020-18-PTB
Deuterium lamp AN5083	CEQ-262-19-005	Nov. 2019	Reference standard DLS-IS-L#01	40022-19-PTB
LED calibration standard white, ACS-570-1 SN: 007357018	LED-19-017	Sep. 2019	Reference standards: - ACS570-1 SN:10008751 - ACS570-1 SN:10008851	47237 PTB 17 47233 PTB 17
LED calibration standard blue, ACS-570-3 SN: 008457018	LED-19-019	Sep. 2019	Reference standards: - ACS570-3 SN:10051551 - ACS570-3 SN:10051651	47238 PTB 17 47234 PTB 17
LED calibration standard green, ACS-570-5 SN: 007557018	LED-19-021	Sep. 2019	Reference standards: - ACS570-5 SN:000457017 - ACS570-5 SN:10052051	47239 PTB 17 47235 PTB 17
LED calibration standard red, ACS-570-7 SN: 007757018	LED-19-023	Sep. 2019	Reference standards: - ACS570-7 SN:000957017 - ACS570-7 SN:10061451	47240 PTB 17 47236 PTB 17

Ambient conditions:

Ambient conditions	
Temperature	23 °C ± 3 °C
Humidity	56% ± 5% relative humidity

Spectrometer settings during luminous flux measurement:

Detector:	PMT3	Averaging:	2
Grating:	1200 g/mm UV	Step width:	1 nm
Wavelength range:	240 – 930 nm	Density filter:	none
Bandpass:	2.5 nm	Cut filter :	yes
Scan speed:	60 ms / nm	PMT high voltage:	3
PMT cooling:	none		

The warm-up time of the spectrometer was 3 hours. The spectrometer SP320-114 and all accessories listed on page 1 were in good order and condition during the entire calibration process.

Certificate No.: CAL-241-20-001

4. Measurement results

a) Wavelength

Table of measured lines:

Nominal value [nm]	Measured value [nm]
253.65	253.65
435.84	435.96
546.08	546.18
632.82	632.93
881.94	881.94

The measurement tolerance of the wavelength measurement is +/-0.2 nm for all wavelengths.

b) Luminous flux of LEDs

For the testing of the measurement accuracy of the spectroradiometer for luminous flux of LEDs the measurement parameters and working standards mentioned in point 3 were used.

Table of results:

LED Serial number	Dominant wavelength [nm] / correlated color temperature [K]	Luminous flux [lm] Nominal value	Luminous flux [lm] Measured value
ACS570-1_007357018	5782	17.72	17.72
ACS570-3_008457018	481.23	7.670	7.849
ACS570-5_007557018	528.10	16.67	16.72
ACS570-7_007757018	622.37	9.485	9.476

The values of dominant wavelength and correlated color temperature are not subject to the test procedure and specified for information only.

Certificate No.: CAL-241-20-001

The measurement uncertainty of the luminous flux depends on the color of the LED:

Color	relative expanded uncertainty
Blue	4.5%
White	3.5%
Green	3.5%
Red	5%

The specified relative expanded uncertainty of measurement corresponds to twice the standard deviation ($k=2$). For a normal distribution, this means that 95 per cent of the measured values lie within the indicated measurement uncertainty interval.

5. Notes

- a) Any mechanical modifications to the tested instrument and the accessories listed on page 1 influence the sensitivity of the instrument.
- b) Instrument Systems certifies that all reference standards and measuring devices have been used within the valid scope of application during the entire test process. The test objects described on page 1 were in good order and condition during the test procedure.
- c) This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.
- d) The results indicated in this certificate refer solely to the instruments tested in the laboratory.
- e) The BaSO₄ coating of the sphere can fatigue due to external influences like dirt, UV-radiation, heat and humidity. These processes can change the reflection of the sphere wall and therefore influence the measurement results.

End of certificate