



Thorlabs GmbH
 Hans-Boeckler-Str. 6
 85221 Dachau
 Germany
 Tel.: +49 8131 5956 0
 Fax.: +49 8131 5956 99

<http://www.thorlabs.com>

Certificate of Calibration

No	19283075349
----	-------------

New Unit
 X Returned Unit

Result as found: FAIL
 Result as left: PASS

Manufacturer Thorlabs GmbH
 Model Number PM100D
 Serial Number P0000219
 Date of Calibration 10-Oct-2019

Calibration Standards

NO	MANUFACTURER	MODEL	SERIAL NO	LAST CAL
1	Keithley Instruments Inc.	Model 2611B	4118325	21.09.2018
2	HEWLETT-PACKARD	34401A	MY45034937	29.02.2019

Thorlabs GmbH does hereby certify that the above mentioned equipment has been calibrated in accordance with our quality management system.

Our Quality Management System is certified according to DIN EN ISO 9001.


The measurement equipment used for calibration is traceable to national standards of the 'EUROMET' members (NPL, PTB, BNM etc.), the US 'NIST' or other national metrological institutions.

Measurements which cannot be traced to national standards can be traced to natural constants, other accepted standards or relational measurements.

Additional documentation concerning traceability of the measurement equipment is available and can be examined upon request.

The certificate of calibration may only be forwarded in complete form without any changes.

The recommended calibration interval is 12 months. The calibration period of this instrument / system begins on the date of receipt by the customer.

Calibrated by  _____

Marlon Becker
 Thorlabs GmbH

Date Received _____

Calibrated Due _____

Test Report - as left

Model PM100D
 Serial No P0000219
 Firmware V2.6.0

Temperature 23°C +/-5°C
 Humidity 35% +/-15%
 Scan CAL_PM100D

Test Date 10-Oct-2019
 Tester Marlon Becker
 Test Result PASS

CURRENT INPUT

AMP RANGE	INPUT CURRENT	DISPLAY (AMP)	UNCERT.	TOL.	ANALOG OUT (VOLT)	UNCERT.	TOL.	RESULT
5 mA	5,0011E-03	5,0011E-03	0,05%	0,20%	2,0106E+00	0,51%	3,00%	PASS
500 µA	5,0004E-04	5,0005E-04	0,05%	0,20%	1,9997E+00	0,02%	3,00%	PASS
50 µA	5,0013E-05	5,0016E-05	0,06%	0,20%	1,9922E+00	0,41%	3,00%	PASS
5 µA	5,0005E-06	5,0004E-06	0,05%	0,20%	1,9982E+00	0,10%	3,00%	PASS
500 nA	5,0013E-07	5,0013E-07	0,05%	0,20%	1,9943E+00	0,31%	3,00%	PASS
50 nA	5,0003E-08	4,9998E-08	0,06%	0,50%	1,9728E+00	1,36%	5,00%	PASS

VOLTAGE INPUT

Pyroelectric Sensor

VOLT RANGE	INPUT VOLTAGE	MEASURED (VOLTS)	UNCERT.	TOL.	ANALOG OUT (VOLT)	UNCERT.	TOL.	RESULT
100 V	3,9999E+01	4,0002E+01	0,06%	0,50%	7,9603E-01	0,49%	3,00%	PASS
10 V	1,0000E+01	1,0002E+01	0,07%	0,50%	1,9848E+00	0,76%	3,00%	PASS
1 V	1,0000E+00	1,0001E+00	0,06%	0,50%	1,9970E+00	0,15%	3,00%	PASS
100 mV	1,0001E-01	1,0007E-01	0,12%	0,50%	1,9926E+00	0,38%	3,00%	PASS

Thermal Sensor

VOLT RANGE	INPUT VOLTAGE	DISPLAY (VOLT)	UNCERT.	TOL.	ANALOG OUT (VOLT)	UNCERT.	TOL.	RESULT
1 V	9,9996E-01	9,9988E-01	0,06%	0,50%	1,9889E+00	0,56%	3,00%	PASS
100 mV	9,9972E-02	9,9983E-02	0,06%	0,50%	1,9889E+00	0,55%	3,00%	PASS
10 mV	1,0000E-02	9,9999E-03	0,05%	0,50%	1,9841E+00	0,80%	3,00%	PASS
1 mV	9,9993E-04	9,9995E-04	0,05%	0,50%	1,9879E+00	0,61%	5,00%	PASS

Test Report - as found

Model PM100D
 Serial No P0000219
 Firmware V2.6.0

Temperature 23°C +/-5°C
 Humidity 35% +/-15%
 Scan CAL_PM100D

Test Date 10-Oct-2019
 Tester Marlon Becker
 Test Result FAIL
 Last Calibration 26-Mar-2009

CURRENT INPUT

AMP RANGE	INPUT CURRENT	DISPLAY (AMP)	UNCERT.	TOL.	ANALOG OUT (VOLT)	UNCERT.	TOL.	RESULT
5 mA	5,0011E-03	4,9995E-03	0,08%	0,20%	2,0105E+00	0,50%	3,00%	PASS
500 µA	5,0004E-04	5,0040E-04	0,12%	0,20%	1,9997E+00	0,02%	3,00%	PASS
50 µA	5,0012E-05	4,9999E-05	0,08%	0,20%	1,9922E+00	0,41%	3,00%	PASS
5 µA	5,0004E-06	5,0114E-06	0,27%	0,20%	1,9982E+00	0,10%	3,00%	FAIL
500 nA	5,0012E-07	5,0012E-07	0,05%	0,20%	1,9944E+00	0,31%	3,00%	PASS
50 nA	5,0003E-08	4,9967E-08	0,12%	0,50%	1,9729E+00	1,36%	5,00%	PASS

VOLTAGE INPUT

Pyroelectric Sensor

VOLT RANGE	INPUT VOLTAGE	MEASURED (VOLTS)	UNCERT.	TOL.	ANALOG OUT (VOLT)	UNCERT.	TOL.	RESULT
100 V	4,0003E+01	3,9891E+01	0,33%	0,50%	7,9602E-01	0,50%	3,00%	PASS
10 V	1,0001E+01	9,9734E+00	0,32%	0,50%	1,9847E+00	0,77%	3,00%	PASS
1 V	1,0000E+00	1,0001E+00	0,06%	0,50%	1,9970E+00	0,15%	3,00%	PASS
100 mV	1,0001E-01	9,9947E-02	0,11%	0,50%	1,9922E+00	0,40%	3,00%	PASS

Thermal Sensor

VOLT RANGE	INPUT VOLTAGE	DISPLAY (VOLT)	UNCERT AINTY	TOL.	ANALOG OUT (VOLT)	UNCERT.	TOL.	RESULT
1 V	9,9995E-01	9,9584E-01	0,46%	0,50%	1,9889E+00	0,55%	3,00%	PASS
100 mV	9,9975E-02	9,9423E-02	0,60%	0,50%	1,9888E+00	0,56%	3,00%	FAIL
10 mV	9,9999E-03	9,8894E-03	1,15%	0,50%	1,9840E+00	0,80%	3,00%	FAIL
1 mV	9,9990E-04	9,6504E-04	3,54%	0,50%	1,9882E+00	0,59%	5,00%	FAIL

Inspection Report

Model PM100D
Serial No P0000219
Software V2.6.0

Temperature 23°C +/-5°C
Humidity 35% +/-15%
Scan CAL_PM100D

Test Date 10-Oct-2019
Tester Marlon Becker
Test Result FAIL
Last Calibration 26-Mar-2009

Incoming Conditions:

Ingoing unit visual in perfect condition:

In need of repair:

Damaged during transportation:

Work Performed:

Firmware update:

Repair:

Modification:

Calibration and verification:

Notes:



Thorlabs GmbH
 Hans-Boeckler-Str. 6
 85221 Dachau
 Germany
 Tel.: +49 8131 5956 0
 Fax.: +49 8131 5956 99

<http://www.thorlabs.com>

Certificate of Calibration

No	19287100346
----	-------------

New Unit
 Returned Unit

Result as found: FAIL
 Result as left: PASS

Manufacturer: Thorlabs GmbH
 Model Number: S122C
 Serial Number: 9032503
 Date of Calibration: 14-Oct-2019

Calibration Standards

NO	MANUFACTURER	MODEL	SERIAL NO	CERTIFICATION	LAST CAL
1	Thorlabs	GM10HS	THO508	73324 19 PTB	3-Apr-2019
2	Hamamatsu	S2281	THO505	73330 19 PTB	16-May-2019
3	Thorlabs	S148C	15050530	NIST 685/0-79-19/1	31-May-2019
4	Thorlabs	S180C	190605400	NIST 685/0-79-19/2	17-Jun-2019

Thorlabs GmbH does hereby certify that the above mentioned equipment has been calibrated in accordance with our quality management system.

Our Quality Management System is certified according to DIN EN ISO 9001.

The measurement equipment used for calibration is traceable to national standards of the 'EUROMET' members (NPL, PTB, BNM etc.), the US 'NIST' or other national metrological institutions.

Measurements which cannot be traced to national standards can be traced to natural constants, other accepted standards or relational measurements.

Additional documentation concerning traceability of the measurement equipment is available and can be examined upon request.

The certificate of calibration may only be forwarded in complete form without any changes.

The recommended calibration interval is 12 months. The calibration period of this instrument / system begins on the date of receipt by the customer.

Calibrated by  Reiner Menning
 Thorlabs GmbH

Date Received _____

Calibrated Due _____


Test Report - as left

Model S122C
 Serial No 9032503
 Scan CAL2

Temperature 25,6°C
 Humidity 44%

Test Date 14-Oct-2019
 Tester Reiner Menning
 Result PASS

λ [nm]	η [A/W]	λ [nm]	η [A/W]	λ [nm]	η [A/W]	λ [nm]	η [A/W]	λ [nm]	η [A/W]	λ [nm]	η [A/W]
700	1,57E-03	885	4,67E-03	1070	6,27E-03	1255	1,70E-02	1440	5,89E-02	1625	5,70E-02
705	1,76E-03	890	4,66E-03	1075	6,34E-03	1260	1,76E-02	1445	6,04E-02	1630	5,52E-02
710	1,94E-03	895	4,75E-03	1080	6,41E-03	1265	1,83E-02	1450	6,19E-02	1635	5,34E-02
715	2,12E-03	900	4,84E-03	1085	6,53E-03	1270	1,90E-02	1455	6,34E-02	1640	5,17E-02
720	2,31E-03	905	4,82E-03	1090	6,68E-03	1275	1,97E-02	1460	6,49E-02	1645	5,02E-02
725	2,51E-03	910	4,78E-03	1095	6,80E-03	1280	2,04E-02	1465	6,64E-02	1650	4,88E-02
730	2,68E-03	915	4,81E-03	1100	6,91E-03	1285	2,12E-02	1470	6,79E-02	1655	4,73E-02
735	2,79E-03	920	4,86E-03	1105	7,04E-03	1290	2,20E-02	1475	6,93E-02	1660	4,58E-02
740	2,87E-03	925	4,87E-03	1110	7,19E-03	1295	2,28E-02	1480	7,07E-02	1665	4,44E-02
745	2,96E-03	930	4,85E-03	1115	7,37E-03	1300	2,37E-02	1485	7,22E-02	1670	4,31E-02
750	3,08E-03	935	4,84E-03	1120	7,55E-03	1305	2,46E-02	1490	7,36E-02	1675	4,18E-02
755	3,23E-03	940	4,85E-03	1125	7,71E-03	1310	2,55E-02	1495	7,49E-02	1680	4,06E-02
760	3,37E-03	945	4,88E-03	1130	7,89E-03	1315	2,65E-02	1500	7,62E-02	1685	3,94E-02
765	3,43E-03	950	4,90E-03	1135	8,10E-03	1320	2,75E-02	1505	7,75E-02	1690	3,83E-02
770	3,49E-03	955	4,90E-03	1140	8,32E-03	1325	2,84E-02	1510	7,88E-02	1695	3,72E-02
775	3,61E-03	960	4,89E-03	1145	8,52E-03	1330	2,94E-02	1515	8,00E-02	1700	3,60E-02
780	3,75E-03	965	4,91E-03	1150	8,72E-03	1335	3,06E-02	1520	8,13E-02	1705	3,48E-02
785	3,89E-03	970	4,95E-03	1155	8,97E-03	1340	3,17E-02	1525	8,25E-02	1710	3,35E-02
790	3,98E-03	975	5,00E-03	1160	9,25E-03	1345	3,29E-02	1530	8,37E-02	1715	3,24E-02
795	4,01E-03	980	5,04E-03	1165	9,52E-03	1350	3,40E-02	1535	8,48E-02	1720	3,14E-02
800	4,04E-03	985	5,08E-03	1170	9,79E-03	1355	3,51E-02	1540	8,57E-02	1725	3,04E-02
805	4,11E-03	990	5,11E-03	1175	1,01E-02	1360	3,63E-02	1545	8,65E-02	1730	2,93E-02
810	4,19E-03	995	5,12E-03	1180	1,03E-02	1365	3,77E-02	1550	8,71E-02	1735	2,83E-02
815	4,26E-03	1000	5,14E-03	1185	1,06E-02	1370	3,90E-02	1555	8,75E-02	1740	2,73E-02
820	4,31E-03	1005	5,20E-03	1190	1,10E-02	1375	4,03E-02	1560	8,75E-02	1745	2,60E-02
825	4,37E-03	1010	5,27E-03	1195	1,13E-02	1380	4,15E-02	1565	8,71E-02	1750	2,47E-02
830	4,44E-03	1015	5,31E-03	1200	1,17E-02	1385	4,29E-02	1570	8,61E-02	1755	2,36E-02
835	4,50E-03	1020	5,37E-03	1205	1,21E-02	1390	4,44E-02	1575	8,46E-02	1760	2,26E-02
840	4,55E-03	1025	5,44E-03	1210	1,25E-02	1395	4,58E-02	1580	8,26E-02	1765	2,16E-02
845	4,55E-03	1030	5,54E-03	1215	1,29E-02	1400	4,71E-02	1585	8,00E-02	1770	2,04E-02
850	4,56E-03	1035	5,63E-03	1220	1,33E-02	1405	4,85E-02	1590	7,69E-02	1775	1,91E-02
855	4,59E-03	1040	5,73E-03	1225	1,38E-02	1410	4,99E-02	1595	7,35E-02	1780	1,80E-02
860	4,63E-03	1045	5,83E-03	1230	1,43E-02	1415	5,13E-02	1600	7,01E-02	1785	1,71E-02
865	4,63E-03	1050	5,90E-03	1235	1,48E-02	1420	5,28E-02	1605	6,71E-02	1790	1,63E-02
870	4,64E-03	1055	5,92E-03	1240	1,53E-02	1425	5,42E-02	1610	6,43E-02	1795	1,53E-02
875	4,68E-03	1060	5,98E-03	1245	1,59E-02	1430	5,58E-02	1615	6,15E-02	1800	1,41E-02
880	4,70E-03	1065	6,12E-03	1250	1,64E-02	1435	5,73E-02	1620	5,90E-02		



Test Report - as found

Model S122C
 Serial No 9032503
 Scan CAL2

Temperature 25,6°C
 Humidity 44%
 Calibrated 25.03.2009

Test Date 14-Oct-2019
 Tester Reiner Menning
 Result FAIL

λ [nm]	η [A/W]	Delta to last Cal	Tol. +/-	Result
700	1,57E-03	-7,7%	6,0%	FAIL
705	1,76E-03	-7,6%	6,0%	FAIL
710	1,94E-03	-7,8%	6,0%	FAIL
715	2,12E-03	-7,6%	6,0%	FAIL
720	2,31E-03	-7,1%	6,0%	FAIL
725	2,51E-03	-5,1%	6,0%	PASS
730	2,68E-03	-4,2%	6,0%	PASS
735	2,79E-03	-5,1%	6,0%	PASS
740	2,87E-03	-7,0%	6,0%	FAIL
745	2,96E-03	-8,2%	6,0%	FAIL
750	3,08E-03	-8,3%	6,0%	FAIL
755	3,23E-03	-7,1%	6,0%	FAIL
760	3,37E-03	-6,5%	6,0%	FAIL
765	3,43E-03	-7,6%	6,0%	FAIL
770	3,49E-03	-8,8%	6,0%	FAIL
775	3,61E-03	-8,3%	6,0%	FAIL
780	3,75E-03	-7,0%	6,0%	FAIL
785	3,89E-03	-5,7%	6,0%	PASS
790	3,98E-03	-5,5%	6,0%	PASS
795	4,01E-03	-6,5%	6,0%	FAIL
800	4,04E-03	-7,6%	6,0%	FAIL
805	4,11E-03	-7,6%	6,0%	FAIL
810	4,19E-03	-7,2%	6,0%	FAIL
815	4,26E-03	-7,1%	6,0%	FAIL
820	4,31E-03	-7,1%	6,0%	FAIL
825	4,37E-03	-6,9%	6,0%	FAIL
830	4,44E-03	-6,4%	6,0%	FAIL
835	4,50E-03	-5,8%	6,0%	PASS
840	4,55E-03	-5,6%	6,0%	PASS
845	4,55E-03	-6,1%	6,0%	FAIL
850	4,56E-03	-6,6%	6,0%	FAIL
855	4,59E-03	-6,3%	6,0%	FAIL
860	4,63E-03	-6,1%	6,0%	FAIL
865	4,63E-03	-6,3%	6,0%	FAIL
870	4,64E-03	-6,5%	6,0%	FAIL
875	4,68E-03	-5,9%	6,0%	PASS
880	4,70E-03	-5,6%	6,0%	PASS
885	4,67E-03	-6,6%	6,0%	FAIL
890	4,66E-03	-7,0%	6,0%	FAIL
895	4,75E-03	-5,4%	6,0%	PASS
900	4,84E-03	-3,7%	6,0%	PASS
905	4,82E-03	-4,1%	6,0%	PASS
910	4,78E-03	-5,0%	6,0%	PASS
915	4,81E-03	-4,4%	6,0%	PASS
920	4,86E-03	-3,5%	6,0%	PASS
925	4,87E-03	-3,5%	6,0%	PASS
930	4,85E-03	-3,9%	6,0%	PASS
935	4,84E-03	-4,3%	6,0%	PASS
940	4,85E-03	-4,2%	6,0%	PASS
945	4,88E-03	-3,9%	6,0%	PASS

λ [nm]	η [A/W]	Delta to last Cal	Tol. +/-	Result
950	4,90E-03	-3,6%	6,0%	PASS
955	4,90E-03	-4,1%	6,0%	PASS
960	4,89E-03	-4,6%	6,0%	PASS
965	4,91E-03	-4,7%	6,0%	PASS
970	4,95E-03	-4,3%	6,0%	PASS
975	5,00E-03	-3,8%	6,0%	PASS
980	5,04E-03	-3,4%	6,0%	PASS
985	5,08E-03	-3,3%	4,0%	PASS
990	5,11E-03	-3,4%	4,0%	PASS
995	5,12E-03	-3,8%	4,0%	PASS
1000	5,14E-03	-4,1%	4,0%	FAIL
1005	5,20E-03	-3,8%	4,0%	PASS
1010	5,27E-03	-3,5%	4,0%	PASS
1015	5,31E-03	-3,5%	4,0%	PASS
1020	5,37E-03	-3,4%	4,0%	PASS
1025	5,44E-03	-3,1%	4,0%	PASS
1030	5,54E-03	-2,5%	4,0%	PASS
1035	5,63E-03	-2,0%	4,0%	PASS
1040	5,73E-03	-1,5%	4,0%	PASS
1045	5,83E-03	-1,0%	4,0%	PASS
1050	5,90E-03	-1,1%	4,0%	PASS
1055	5,92E-03	-2,1%	4,0%	PASS
1060	5,98E-03	-2,6%	4,0%	PASS
1065	6,12E-03	-1,7%	4,0%	PASS
1070	6,27E-03	-0,8%	4,0%	PASS
1075	6,34E-03	-1,3%	4,0%	PASS
1080	6,41E-03	-2,0%	4,0%	PASS
1085	6,53E-03	-1,8%	4,0%	PASS
1090	6,68E-03	-1,2%	4,0%	PASS
1095	6,80E-03	-1,5%	4,0%	PASS
1100	6,91E-03	-1,8%	4,0%	PASS
1105	7,04E-03	-1,9%	4,0%	PASS
1110	7,19E-03	-1,7%	4,0%	PASS
1115	7,37E-03	-1,3%	4,0%	PASS
1120	7,55E-03	-1,0%	4,0%	PASS
1125	7,71E-03	-1,2%	4,0%	PASS
1130	7,89E-03	-1,2%	4,0%	PASS
1135	8,10E-03	-1,0%	4,0%	PASS
1140	8,32E-03	-0,7%	4,0%	PASS
1145	8,52E-03	-1,0%	4,0%	PASS
1150	8,72E-03	-1,1%	4,0%	PASS
1155	8,97E-03	-0,8%	4,0%	PASS
1160	9,25E-03	-0,2%	4,0%	PASS
1165	9,52E-03	-0,3%	4,0%	PASS
1170	9,79E-03	-0,3%	4,0%	PASS
1175	1,01E-02	-0,3%	4,0%	PASS
1180	1,03E-02	-0,2%	4,0%	PASS
1185	1,06E-02	-0,6%	4,0%	PASS
1190	1,10E-02	-0,6%	4,0%	PASS
1195	1,13E-02	-0,2%	4,0%	PASS

Test Report - as found

Model S122C
 Serial No 9032503
 Scan CAL2

Temperature 25,6°C
 Humidity 44%
 Calibrated 25.03.2009

Test Date 14-Oct-2019
 Tester Reiner Menning
 Result FAIL

λ [nm]	η [A/W]	Delta to last Cal	Tol. +/-	Result
1200	1,17E-02	0,2%	4,0%	PASS
1205	1,21E-02	-0,1%	4,0%	PASS
1210	1,25E-02	-0,3%	4,0%	PASS
1215	1,29E-02	-0,4%	4,0%	PASS
1220	1,33E-02	-0,2%	4,0%	PASS
1225	1,38E-02	-0,1%	4,0%	PASS
1230	1,43E-02	0,0%	4,0%	PASS
1235	1,48E-02	-0,2%	4,0%	PASS
1240	1,53E-02	-0,3%	4,0%	PASS
1245	1,59E-02	-0,4%	4,0%	PASS
1250	1,64E-02	-0,2%	4,0%	PASS
1255	1,70E-02	-0,4%	4,0%	PASS
1260	1,76E-02	-0,5%	4,0%	PASS
1265	1,83E-02	-0,2%	4,0%	PASS
1270	1,90E-02	0,3%	4,0%	PASS
1275	1,97E-02	-0,1%	4,0%	PASS
1280	2,04E-02	-0,3%	4,0%	PASS
1285	2,12E-02	0,0%	4,0%	PASS
1290	2,20E-02	0,4%	4,0%	PASS
1295	2,28E-02	0,0%	4,0%	PASS
1300	2,37E-02	-0,2%	4,0%	PASS
1305	2,46E-02	0,1%	4,0%	PASS
1310	2,55E-02	0,6%	4,0%	PASS
1315	2,65E-02	0,4%	4,0%	PASS
1320	2,75E-02	0,1%	4,0%	PASS
1325	2,84E-02	-0,1%	4,0%	PASS
1330	2,94E-02	-0,1%	4,0%	PASS
1335	3,06E-02	-0,1%	4,0%	PASS
1340	3,17E-02	0,1%	4,0%	PASS
1345	3,29E-02	0,3%	4,0%	PASS
1350	3,40E-02	0,3%	4,0%	PASS
1355	3,51E-02	0,8%	4,0%	PASS
1360	3,63E-02	1,4%	4,0%	PASS
1365	3,77E-02	1,3%	4,0%	PASS
1370	3,90E-02	1,3%	4,0%	PASS
1375	4,03E-02	1,4%	4,0%	PASS
1380	4,15E-02	1,6%	4,0%	PASS
1385	4,29E-02	1,0%	4,0%	PASS
1390	4,44E-02	0,5%	4,0%	PASS
1395	4,58E-02	0,9%	4,0%	PASS
1400	4,71E-02	1,1%	4,0%	PASS
1405	4,85E-02	0,9%	4,0%	PASS
1410	4,99E-02	0,8%	4,0%	PASS
1415	5,13E-02	0,8%	4,0%	PASS
1420	5,28E-02	0,9%	4,0%	PASS
1425	5,42E-02	0,6%	4,0%	PASS
1430	5,58E-02	0,4%	4,0%	PASS
1435	5,73E-02	0,5%	4,0%	PASS
1440	5,89E-02	0,6%	4,0%	PASS
1445	6,04E-02	0,5%	4,0%	PASS

λ [nm]	η [A/W]	Delta to last Cal	Tol. +/-	Result
1450	6,19E-02	0,4%	4,0%	PASS
1455	6,34E-02	0,3%	4,0%	PASS
1460	6,49E-02	0,3%	4,0%	PASS
1465	6,64E-02	0,5%	4,0%	PASS
1470	6,79E-02	0,7%	4,0%	PASS
1475	6,93E-02	0,5%	4,0%	PASS
1480	7,07E-02	0,2%	4,0%	PASS
1485	7,22E-02	0,4%	4,0%	PASS
1490	7,36E-02	0,5%	4,0%	PASS
1495	7,49E-02	0,3%	4,0%	PASS
1500	7,62E-02	0,1%	4,0%	PASS
1505	7,75E-02	0,2%	4,0%	PASS
1510	7,88E-02	0,3%	4,0%	PASS
1515	8,00E-02	0,2%	4,0%	PASS
1520	8,13E-02	0,2%	4,0%	PASS
1525	8,25E-02	0,3%	4,0%	PASS
1530	8,37E-02	0,5%	4,0%	PASS
1535	8,48E-02	0,4%	4,0%	PASS
1540	8,57E-02	0,3%	4,0%	PASS
1545	8,65E-02	0,2%	4,0%	PASS
1550	8,71E-02	0,0%	4,0%	PASS
1555	8,75E-02	0,2%	4,0%	PASS
1560	8,75E-02	-0,1%	4,0%	PASS
1565	8,71E-02	-0,1%	4,0%	PASS
1570	8,61E-02	-0,6%	4,0%	PASS
1575	8,46E-02	-0,2%	4,0%	PASS
1580	8,26E-02	-0,4%	4,0%	PASS
1585	8,00E-02	-0,1%	4,0%	PASS
1590	7,69E-02	-0,4%	4,0%	PASS
1595	7,35E-02	-0,4%	4,0%	PASS
1600	7,01E-02	-0,3%	4,0%	PASS
1610	6,43E-02	0,2%	4,0%	PASS
1620	5,90E-02	-0,5%	4,0%	PASS
1630	5,52E-02	0,0%	4,0%	PASS
1640	5,17E-02	-0,3%	4,0%	PASS
1650	4,88E-02	0,1%	4,0%	PASS
1660	4,58E-02	-0,7%	7,0%	PASS
1670	4,31E-02	-0,9%	7,0%	PASS
1680	4,06E-02	-0,6%	7,0%	PASS
1690	3,83E-02	-0,4%	7,0%	PASS
1700	3,60E-02	0,1%	7,0%	PASS
1710	3,35E-02	-0,5%	7,0%	PASS
1720	3,14E-02	0,0%	7,0%	PASS
1730	2,93E-02	0,3%	7,0%	PASS
1740	2,73E-02	0,9%	7,0%	PASS
1750	2,47E-02	-0,6%	7,0%	PASS
1760	2,26E-02	0,0%	7,0%	PASS
1770	2,04E-02	0,5%	7,0%	PASS
1785	1,71E-02	0,8%	7,0%	PASS
1800	1,41E-02	1,9%	7,0%	PASS

Inspection Report

Model S122C
Serial No 9032503

Test Date 14-Oct-2019
Tester Reiner Menning

Incoming Conditions:

Ingoing unit visual in perfect condition: YES

In need of repair: NO

Damaged during transportation: NO

Work Performed:

Cleaning: NO

Repair: NO

Modification: NO

Calibration and Verification: YES

Notes:

This page intentionally left blank